

Notice of Allowability

Application No.

10/065,893

Examiner

John F. Ramirez

Applicant(s)

JENKINS, JOHN H.

Art Unit

3737

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to 02/22/07.
2. ☒ The allowed claim(s) is/are 3, 4, 6-17, and 37-44.
3. ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☐ All b) ☐ Some* c) ☐ None of the:
 1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).


* Certified copies not received: _____.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.
THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
 5. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____.
 - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

1. ☐ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☐ Information Disclosure Statements (PTO/SB/08),
Paper No./Mail Date _____
4. ☐ Examiner's Comment Regarding Requirement for Deposit
of Biological Material
5. ☐ Notice of Informal Patent Application
6. ☐ Interview Summary (PTO-413),
Paper No./Mail Date _____
7. ☐ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other _____


BRIAN L. CASLER
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 3700

Allowable Subject Matter

The following is an examiner's statement of reasons for allowance:

The prior art of record does not disclose or suggest, *inter alia*, a method comprising the following steps: (a) acquiring frames of imaging data representing a succession of images of a patient's heart during a study using an imaging system; (b) automatically time stamping each acquired frame with respective times generated by a first clock; (c) automatically calculating respective offsets of said first clock relative to a reference clock for each time stamp associated with said acquired frames; (d) storing said acquired frames and said time stamps and said offsets associated with said acquired frames; (e) acquiring physiologic data from said patient during said study using a hemodynamic monitoring system; (f) automatically time stamping data in said acquired physiologic data representing one or more predetermined cardiac events with respective times generated by a second clock; (g) automatically calculating respective offsets of said second clock relative to said reference clock for each time stamp associated with said physiologic data; (h) storing said acquired physiologic data and said time stamps and said offsets associated with said acquired physiologic data, (i) computing which one of said acquired frames was acquired in substantial synchronism with a first predetermined cardiac event based on said time stamps and said offsets; and (j) associating a first acquired frame identified by step (i) with said first predetermined cardiac event, further comprising the step of displaying said first acquired frame of imaging data concurrently with acquired physiologic data having a visual indicator indicating a physiologic datum corresponding to said first predetermined cardiac event.

Art Unit: 3737

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to John F. Ramirez whose telephone number is (571) 272-8685. The examiner can normally be reached on (Mon-Fri) 7:30 - 4:00 p.m.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Brian L. Casler can be reached on (571) 272-4956. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

JFR

BRIAN L. CASLER
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 3700